



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE
SCHEME

EQAS YEARLY SUMMARY - 2022



PC-1024

Lab Name **SRI MANAKULA VINAYAGAR MEDICAL COLLEGE&HOSPITAL**

Lab No **1871**

Constituent Group **Chemistry I**

PT item **Lyophilized human serum based**

Date of Report **20-02-2023**
Published

Group I : Laboratories who reported 10 to 12 months during 2022

Group II : Laboratories who reported 6 to 9 months during 2022

Group III : Laboratories who reported less than 6 months during 2022

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	GLUCOSE	Group I	3343	0.71	A
2	UREA	Group I	3306	0.31	A
3	CREATININE	Group I	3301	0.43	A
4	T.BILIRUBIN	Group I	3233	0.68	A
5	T-PROTEIN	Group I	3244	0.59	A
6	ALBUMIN	Group I	3202	0.36	A
7	CALCIUM	Group I	2703	0.59	A
8	PHOSPHORUS	Group I	1863	0.84	A
9	URIC ACID	Group I	3166	0.95	A
10	CHOLESTEROL	Group I	3242	0.27	A
11	TRIGLYCERIDE	Group I	3139	0.46	A
12	HDL	Group I	2687	0.61	A
13	SODIUM	Group I	2542	0.41	A
14	POTASSIUM	Group I	2531	0.63	A
15	CHLORIDE	Group I	2021	0.62	A
16	AST	Group I	2804	0.83	A
17	ALT	Group I	2803	0.64	A
18	ALP	Group I	3136	0.54	A
19	AMYLASE	Group I	1881	0.75	A
20	MAGNESIUM	Group I	786	1.18	B

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks
0.00 - 1.00	A	Excellent
1.01 - 2.00	B	Good
2.01 - 2.99	C	Satisfactory
Beyond ± 3.00	D	Unacceptable

LAB ADDRESS
SRI MANAKULA VINAYAGAR MEDICAL COLLEGE&HOSPITAL
DEPT.BIOCHEMISTRY, KALITHERTHALKUPPAM, MADAGADIPET
PONDICHERRY
PUDUCHERI605107

Pamela Christudoss



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE
SCHEME

EQAS YEARLY SUMMARY - 2022



PC-1024

Lab Name **SRI MANAKULA VINAYAGAR MEDICAL COLLEGE&HOSPITAL** Lab No **1871**
Constituent Group **Thyroid Hormones & Cortisol**
PT item **Lyophilized human serum based** Date of Report **20-02-2023**
Published

Group I : Laboratories who reported 10 to 12 months during 2022

Group II : Laboratories who reported 6 to 9 months during 2022

Group III : Laboratories who reported less than 6 months during 2022

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	FREE THYROXINE	Group I	876	0.94	A
2	TSH	Group I	1342	0.89	A
3	CORTISOL	Group I	247	0.55	A

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks
0.00 - 1.00	A	Excellent
1.01 - 2.00	B	Good
2.01 - 2.99	C	Satisfactory
Beyond ± 3.00	D	Unacceptable

LAB ADDRESS

SRI MANAKULA VINAYAGAR MEDICAL COLLEGE&HOSPITAL
DEPT.BIOCHEMISTRY, KALITHERTHALKUPPAM, MADAGADIPET
PONDICHERRY
PUDUCHERI605107

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact number:0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College,
Vellore

Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE
SCHEME

EQAS YEARLY SUMMARY - 2022



PC-1024

Lab Name **SRI MANAKULA VINAYAGAR MEDICAL COLLEGE&HOSPITAL**

Lab No **1871**

Constituent Group **Tumor Markers**

PT item **Lyophilized human serum based**

Date of Report **20-02-2023**
Published

Group I : Laboratories who reported 5 to 6 months during 2022

Group II : Laboratories who reported 3 to 4 months during 2022

Group III : Laboratories who reported less than 3 months during 2022

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	PSA (TOTAL)	Group I	169	0.90	A

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks
0.00 - 1.00	A	Excellent
1.01 - 2.00	B	Good
2.01 - 2.99	C	Satisfactory
Beyond ± 3.00	D	Unacceptable

LAB ADDRESS

SRI MANAKULA VINAYAGAR MEDICAL COLLEGE&HOSPITAL
DEPT.BIOCHEMISTRY, KALITHERTHALKUPPAM, MADAGADIPET
PONDICHERRY
PUDUCHERI605107

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact number:0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College,
Vellore

Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE
SCHEME

EQAS YEARLY SUMMARY - 2022



PC-1024

Lab Name **SRI MANAKULA VINAYAGAR MEDICAL COLLEGE&HOSPITAL**
Constituent Group **Reproductive Hormones**
PT item **Lyophilized human serum based**

Lab No **1871**

Date of Report **20-02-2023**
Published

Group I : Laboratories who reported 10 to 12 months during 2022

Group II : Laboratories who reported 6 to 9 months during 2022

Group III : Laboratories who reported less than 6 months during 2022

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	LH	Group I	573	0.82	A
2	FSH	Group I	582	0.73	A
3	PROLACTIN	Group I	589	0.69	A
4	TOTAL BETA HCG	Group I	484	1.12	B
5	ESTRADIOL	Group I	292	0.46	A
6	PROGESTERONE	Group I	233	0.56	A

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks
0.00 - 1.00	A	Excellent
1.01 - 2.00	B	Good
2.01 - 2.99	C	Satisfactory
Beyond ± 3.00	D	Unacceptable

LAB ADDRESS

SRI MANAKULA VINAYAGAR MEDICAL COLLEGE&HOSPITAL
DEPT.BIOCHEMISTRY, KALITHERTHALKUPPAM, MADAGADIPET
PONDICHERRY
PUDUCHERI605107

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College,
Vellore

Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE
SCHEME

EQAS YEARLY SUMMARY - 2022



PC-1024

Lab Name **SRI MANAKULA VINAYAGAR MEDICAL COLLEGE&HOSPITAL** Lab No **1871**
Constituent Group **HbA1c**
PT item **Lyophilized human whole blood based** Date of Report Published **20-02-2023**

Group I : Laboratories who reported 10 to 12 months during 2022

Group II : Laboratories who reported 6 to 9 months during 2022

Group III : Laboratories who reported less than 6 months during 2022

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	HbA1c	Group I	1541	1.34	B

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks
0.00 - 1.00	A	Excellent
1.01 - 2.00	B	Good
2.01 - 2.99	C	Satisfactory
Beyond ± 3.00	D	Unacceptable

LAB ADDRESS

SRI MANAKULA VINAYAGAR MEDICAL COLLEGE&HOSPITAL
DEPT.BIOCHEMISTRY, KALITHERTHALKUPPAM, MADAGADIPET
PONDICHERRY
PUDUCHERI605107

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College,
Vellore

Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components

***** End of Report *****